

August 15, 2023

Mr. Felipe Ramos
Analytical Industries Inc
2855 Metropolitan Place,
Pomona,
CA 91767
United States of America

Subject: Gas Analyzers Portable, Online and Loop Powered
Listing Number E115647; MET Project Number PRJ15816

Safety Standards:	UL 60079-0	CSA C22.2 No. 60079-0
	UL 60079-11	CSA C22.2 No. 60079-11
	UL 60079-1	CSA C22.2 NO. 60079-1
	UL 1203	CSA C22.2 NO. 30
	UL 61010-1	CSA C22.2 No. 61010-1

Dear Mr. Ramos:

Congratulations on successfully completing the MET Mark Certification process for the Gas Analyzers Portable, Online and Loop Powered. Analytical Industries Inc may begin to apply the MET Mark on the previously identified product at this time in accordance with the MET Mark Applicant Contract. The report covering the above stated product is forthcoming.

Follow-up inspections are conducted unannounced and quarterly to assure the Certified product is identical to the product evaluated. Your initial inspection will occur sometime within the first three months after the approval is issued. However, the first inspection will be an announced inspection. Someone from our follow-up services will contact you several weeks ahead of time to coordinate the first inspection.

Thank you for the opportunity to perform this service for Analytical Industries Inc. We look forward to future opportunities with your company.

Sincerely,

MET LABORATORIES, INC.



D. Benstead
Technical Manager,
Eurofins E&E CML Ltd.



Canadian Certification has been granted under a System 3 program as defined in ISO/IEC 17067.

NRTL