



505 - 6th Street, Suite 200  
New Westminster, BC V3L 0E1  
Toll Free: 1-866-566-SAFE  
Fax: (778) 396 - 2064  
[www.safetyauthority.ca](http://www.safetyauthority.ca)

PRESSURE VESSEL ENGINEERING  
B-120 RANDALL DRIVE  
WATERLOO ON N2V 1C6

Date: July 14, 2014  
Account #: 45363  
Journal #: 59793  
Our File #: 5497411

**Re: Application for Design Registration**

The design, as detailed in your, PVE-7853, for a Fitting is accepted for registration as follows:

**Registered To: MICHELL INSTRUMENTS LIMITED CRN: 0F04115.61**

**MDMT: -20 deg F      MAWT: 158 deg F      MAWP: 4800 psig**

**Drawing #: EX90566**

**Drawing Revision: 01**

**Reviewer's Notes:**

- 1) This registration is for an Easidew Pro XP Sensor Body.
- 2) This registration expires on April 29, 2024.
- 3) This registration is valid until the indicated expiry date only if the Manufacturer maintains a valid quality management system approved by an acceptable third-party agency until that date. Should the approval of the quality management system lapse before the expiry date indicated above, this registration shall become void.

Contact me if you have any questions. The invoice for registration will be forwarded under separate cover.

HEYE KLAVER  
(778) 396-2030  
[Heye.Klaver@safetyauthority.ca](mailto:Heye.Klaver@safetyauthority.ca)  
Design Engineer

cc: