



QPS Evaluation Services Inc
Testing, Certification and Field Evaluation Body
Accredited in Canada, the USA, and Internationally

File
LR1551

CERTIFICATE OF COMPLIANCE
(ISO TYPE 3 CERTIFICATION SYSTEM)

Issued to	Michell Instrument UK
Address	48 Lancaster Way Business Park, Ely Cambridgeshire, CB6 3NW, United Kingdom
Project Number	LR1551-3
Product	Trace Moisture Analyzer
Model Number	QMA401
Ratings	85-264Vac, 50/60Hz, 150VA
Applicable Standards	<ul style="list-style-type: none"> • CAN/CSA C22.2 No. 61010-1 - Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 1: General requirements - Third Edition; Update No. 1: July 2015; Update No. 2: April 2016; Amd 1: November 2018 • UL 61010-1 - UL Standard for Safety Electrical Equipment for Measurement, Control, and Laboratory Use; Part 1: General Requirements - Third Edition; Including Revisions through November 21, 2018
Factory/Manufacturing Location	Same as applicant

Statement of Compliance: The product(s) identified in this Certificate and described in the Report covered under the above referenced project number have been investigated and found to be in compliance with the relevant requirements of the above referenced standard(s). As such, they are eligible to bear the QPS Certification Mark shown below, in accordance with the provisions of QPS's Service Agreement.



Issued By: Simon Hodson

Signature: *Simon Hodson*

Date: October 8, 2020