



**QPS Evaluation Services Inc**  
 Testing, Certification and Field Evaluation Body  
 Accredited in Canada, the USA, and Internationally

File
LR1551

**CERTIFICATE OF COMPLIANCE**  
 (ISO TYPE 3 CERTIFICATION SYSTEM)

Issued to	Michell Instrument UK
Address	48 Lancaster Way Business Park Ely, Cambridgeshire, CB6 3NW, United Kingdom
Project Number	LR1551-2
Product	Dew-Point Transmitter
Model Number	SF82
Ratings	5-28Vdc, 4-20mA, 23mA Max
Applicable Standards	<ul style="list-style-type: none"> <li>• CAN/CSA C22.2 No. 61010-1 - Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 1: General requirements - Third Edition; Update No. 1: July 2015; Update No. 2: April 2016 – (R2017)</li> <li>• UL 61010-1 - UL Standard for Safety Electrical Equipment For Measurement, Control, and Laboratory Use; Part 1: General Requirements - Third Edition; Including Revisions through November 21, 2018</li> </ul>
Factory/Manufacturing Location	Same as applicant

Statement of Compliance: The product(s) identified in this Certificate and described in the Report covered under the above referenced project number have been investigated and found to be in compliance with the relevant requirements of the above referenced standard(s). As such, they are eligible to bear the QPS Certification Mark shown below, in accordance with the provisions of QPS's Service Agreement.



Issued By: Simon Hodson

Signature: *Simon Hodson*

Date: August 19, 2019