



**QPS Evaluation Services Inc**  
**Testing, Certification and Field Evaluation Body**  
**Accredited in Canada, the USA, and Internationally**

File  
 LR1507-1

**CERTIFICATE OF COMPLIANCE**  
**(ISO TYPE 3 CERTIFICATION SYSTEM)**

Issued to	Michell Instruments Ltd
Address	Unit 48, Lancaster way Business Park, Ely, Cambs, CB6 3NW, UK
Project Number	LR1507
Product	Tuneable Diode Laser
Model Number	OptiPEAK TDL600
Ratings/Markings	Class I, Division 1, Groups A, B, C and D T5 Ta=-20°C to +60°C, IP66 90-260 VAC, 50/60 Hz, 180 W
Applicable Standards	C22.2 No. 30 -1986, C22.2 No. 60079-0:15, C22.2 No. 60079-11:14, CSA C22.2 No. 61010-1:12, CSA C22. No. 60079-28:16, UL1203 5th ed., UL 913 8th ed., UL 60079-28 2nd ed., UL61010-1 3rd. ed.
Factory/Manufacturing Location	Same as above

**Statement of Compliance:** The product(s) identified in this Certificate and described in the Report covered under the above referenced project number have been investigated and found to be in compliance with the relevant requirements of the above referenced standard(s). As such, they are eligible to bear the QPS Certification Mark shown below, in accordance with the provisions of QPS's Service Agreement.



Issued By: Dave Adams

Signature:

Date: November 30, 2018