

Manufacturer: **Michell Instruments Limited**
48 Lancaster Way Business Park
Ely, Cambridgeshire
CB6 3NW. UK.



On behalf of the above named company, I declare that under our sole responsibility, on the date that the equipment accompanied by this declaration is placed on the market, it conforms with all technical and regulatory requirements of the listed EU Directives listed below

DG2 & DG3 Dew-point Generator

2014/30/EU

EMC Directive

**2011/65/EU-2015/863/EU
Directive (RoHS3)**

Restriction of Hazardous Substances

2011/35/EU

Low Voltage Directive

2014/68/EU


PE Directive

This product and sample systems & accessories that may be supplied with does not bear CE marking for the Pressure Equipment Directive, and are supplied in accordance with Article 4, paragraph 3 of 2014/68/EU by using SEP (sound engineering practice) in the design and manufacturer and are provided with adequate instructions for use.

and has been designed to be in conformance with the relevant sections of the following standards or other normative documents.

EN61326-1:2013 Electrical equipment for measurement, control and laboratory use – EMC requirements –Class A (emissions) and Industrial Locations (immunity).

EN61010-1:2010 Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements



Andrew M.V. Stokes, Technical Director
June 2022
Ely. UK

Manufacturer: **Michell Instruments Limited**
48 Lancaster Way Business Park
Ely, Cambridgeshire
CB6 3NW. UK.



On behalf of the above named company, we the manufacturer declare that under our sole responsibility, on the date that the equipment accompanied by this declaration is placed on the market, the equipment conforms with all technical and regulatory requirements of the relevant sections of standards and/or other normative documents of the listed UKCA Regulations.

DG2 & DG3 Dew-point Generator

The Electromagnetic Compatibility Regulations 2016.

UK Statutory Instrument (S.I. 2016 No.1091).

Is in conformity with the following Standard(s) or Normative Document(s):
EN61326-1:2013 *Electrical equipment for measurement, control and laboratory use - EMC requirements. Class B (emissions) and Industrial Locations (immunity).*

The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012.

UK Statutory Instrument (S.I. 2012 No.3032).

The Electrical Equipment Safety Regulation 2016.

UK Statutory Instrument (S.I. 2016 1101).

BS/EN 61010-1:2010/A1:2019 Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements

The Pressure Equipment (Safety) Regulations 2016.

UK Statutory Instrument (S.I. 2016 No.1105).

This product and sample systems & accessories that may be supplied with it does not bear UKCA marking for the Pressure Equipment (Safety) regulation, and are supplied in accordance with Article 4, paragraph 3 of 2014/68/EU by using SEP (sound engineering practice) in the design and manufacturer and are provided with adequate instructions for use.



Andrew M.V. Stokes, Technical Director
Ely, UK
June 2022