

Manufacturer: **Michell Instruments Limited**  
Address: **48 Lancaster Way Business Park, Ely, Cambridgeshire, CB6 3NW, UK**



declare, under our sole responsibility, on the date that it is placed on the market, the product(s):

## OptiCal Humidity Calibrator

to which this declaration relates, is/are in conformity with the latest technical requirements of the following EU directives and is in conformance with the relevant sections of the harmonised EN standard(s):

### **Low Voltage (LVD) - 2014/35/EU**

EN 61010-1:2010+A1:2019

### **Electromagnetic Compatibility (EMC) - 2014/30/EU**

EN 61326-1:2013 – *Class B (emissions) and Industrial Locations (immunity)*

### **Restriction of the Use of Certain Hazardous Substances (RoHS2) - 2011/65/EU+(EU)2015/863**

EN IEC 63000:2018

Authorised Representative: **PST – PO Box 13386, Navan, Co Meath, Ireland**



---

Jim Macfarlane, Managing Director  
May 2025,  
Ely, UK

Manufacturer: **Michell Instruments Limited**  
Address: **48 Lancaster Way Business Park, Ely, Cambridgeshire, CB6 3NW, UK**



declare, under our sole responsibility, on the date that it is placed on the market, the product(s):

## OptiCal Humidity Calibrator

to which this declaration relates, is/are in conformity with the latest technical requirements of the following UK regulations and is in conformance with the relevant sections of the designated EN standard(s):

**The Electrical Equipment Safety Regulation 2016 (S.I. 2016/1101)**

EN 61010-1:2010+A1:2019

**The Electromagnetic Compatibility Regulations 2016 (S.I. 2016/1091)**

EN 61326-1:2013 – *Class B (emissions) and Industrial Locations (immunity)*

**The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012 (S.I. 2012/3032)**

EN IEC 63000:2018



---

Jim Macfarlane, Managing Director  
May 2025,  
Ely, UK