

Manufacturer: **Michell Instruments Limited**
Address: **48 Lancaster Way Business Park, Ely, Cambridgeshire, CB6 3NW, UK**



declare, under our sole responsibility, on the date that it is placed on the market, the product(s):

S8000 HT Chilled Mirror Hygrometer

S8K431 – USB + Ethernet

S8K432 – USB + RS485

to which this declaration relates, is/are in conformity with the latest technical requirements of the following EU directives and is in conformance with the relevant sections of the harmonised EN standard(s):

Low Voltage (LVD) - 2014/35/EU

EN 61010-1:2010+A1:2019+AC:2019-04

EN IEC 61010-2-010:2020

Electromagnetic Compatibility (EMC) - 2014/30/EU

EN 61326-1:2021 – *Class A (Emissions) and Industrial Locations (Immunity)*

Restriction of the Use of Certain Hazardous Substances (RoHS2) - 2011/65/EU+(EU)2015/863

EN IEC 63000:2018

Authorised Representative: **PST – PO Box 13386, Navan, Co Meath, Ireland**



Stuart Toms, Senior Engineering Director, Product Excellence
March 2026,
Ely, UK

Manufacturer: **Michell Instruments Limited**
Address: **48 Lancaster Way Business Park, Ely, Cambridgeshire, CB6 3NW, UK**



declare, under our sole responsibility, on the date that it is placed on the market, the product(s):

S8000 HT Chilled Mirror Hygrometer

S8K431 – USB + Ethernet

S8K432 – USB + RS485

to which this declaration relates, is/are in conformity with the latest technical requirements of the following UK regulations and is in conformance with the relevant sections of the designated EN standard(s):

The Electrical Equipment Safety Regulation 2016 (S.I. 2016/1101)

EN 61010-1:2010+A1:2019+AC:2019-04

EN IEC 61010-2-010:2020

The Electromagnetic Compatibility Regulations 2016 (S.I. 2016/1091)

EN 61326-1:2021 - *Class A (emissions) and Industrial Locations (immunity)*

The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012 (S.I. 2012/3032)

EN IEC 63000:2018



Stuart Toms, Senior Engineering Director, Product Excellence
March 2026,
Ely, UK