

EU Declaration of Conformity



Manufacturer: **Michell Instruments B.V.**
Kombraak 11
4906 CR Oosterhout
The Netherlands



On behalf of the above named company, I declare that under our sole responsibility, on the date that the equipment accompanied by this declaration is placed on the market, it conforms with all technical and regulatory requirements of the listed EU Directives listed below and has been designed to be in conformance with the relevant sections of the following standards or other normative documents.

SF52 Dew-Point Transmitter

Electromagnetic Compatibility (EMC) - 2014/30/EU

EN 61326-1:2013 & EN 61326-1:2021

Electrical equipment for measurement, control, and laboratory use – EMC requirements – Class A (emissions) and Industrial Locations (immunity).

Restriction of the Use of Certain Hazardous Substances (RoHS) - 2011/65/EU+(EU)2015/863 (RoHS3)

Low Voltage (LVD) - 2014/35/EU

EN 60529:1992+A2:2013

Degrees of Protection Provided by Enclosures (IP Code). Ingress Rating equivalent to IP66.

EN 61010-1:2010+A1:2019

Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements.

Pressure Equipment (PED) - 2014/68/EU

This product and sample systems & accessories that may be supplied with does not bear CE marking for the Pressure Equipment Directive, and are supplied in accordance with Article 4, paragraph 3 of 2014/68/EU by using SEP (sound engineering practice) in the design and manufacturer and are provided with adequate instructions for use.

A handwritten signature in blue ink, appearing to read 'R. Pouw', written over a horizontal line.

Robert-Jan Pouw, RH Business Manager.
On behalf of,
Kees Roos, Managing Director, Michell Instruments Benelux B.V

November 2022
Oosterhout, The Netherlands

Manufacturer: **Michell Instruments B.V.**
Kombraak 11
4906 CR Oosterhout
The Netherlands



On behalf of the above named company, I declare that under our sole responsibility, on the date that the equipment accompanied by this declaration is placed on the market, it conforms with all technical and regulatory requirements of the listed UK Directives listed below and has been designed to be in conformance with the relevant sections of the following standards or other normative documents.

SF52 Dew-Point Transmitter

The Electromagnetic Compatibility Regulations 2016. **UK Statutory Instrument (S.I. 2016 No.1091)**

EN 61326-1:2013 & EN 61326-1:2021

Electrical equipment for measurement, control, and laboratory use – EMC requirements – Class A (emissions) and Industrial Locations (immunity).

The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012. **UK Statutory Instrument (S.I. 2012 No.3032)**

The Electrical Equipment Safety Regulation 2016. **UK Statutory Instrument (S.I. 2016 1101)**

EN 60529:1992+A2:2013

Degrees of Protection Provided by Enclosures (IP Code). Ingress Rating equivalent to IP66.

EN 61010-1:2010+A1:2019

Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements.

The Pressure Equipment (Safety) Regulations 2016. **UK Statutory Instrument (S.I. 2016 No.1105)**

This product and sample systems & accessories that it may be supplied with does not bear UKCA marking for the Pressure Equipment Directive, and are supplied in accordance with part 1, paragraph 8 of S.I. 2016/1105 by using SEP (sound engineering practice) in the design and manufacture and are provided with adequate instructions for use.



Robert-Jan Pouw, RH Business Manager.
On behalf of,
Kees Roos, Managing Director, Michell Instruments Benelux B.V

November 2022
Oosterhout, The Netherlands